K003547



TERANG NUSA Sdn Bhd

DEC 21 2000

510(k) Summary NUGARD DG

510(k) Summary

Submitter Name	Terang Nusa Sdn Bhd
Submitter Address	1, Jalan 8, Pengkalan Chepa 2 Industrial Zone, 16100 Kota Bharu,
	Kelantan, Malaysia.
Submitter Telephone	+60 9 7735133
Submitter Fax	+60 9 7737755
Submitter e-mail	tnsb@po.jaring.my
Contact Person	LOW, Chin Guan
Date of preparation	20 Oct 2000
Trade Name	NUGARD DG, Green-Mint
Common Name	Latex Examination Glove, Green-Mint, Contain less than 100 microgram/gram of water extractable protein.
Classification	Patient Examination Glove
Legally marketed device to which substantial equivalence is being claimed.	The NUGARD DG, examination glove described in this 510(k) is substantially equivalent to the NUGARD latex examination glove that is currently marketed.
Description of device	NUGARD DG meet the requirement for examination gloves described by the American Standard for Testing and Material ASTM D3578(00), green in color, mint scented and prepowdered. Sizes available is from XS – XL

K00 3547



TERANG NUSA Sdn Bhd

510(k) Summary NUGARD DG

These examination gloves are to be worn by
healthcare workers or similar personnel during
work to prevent cross contamination between
the user and the patient.
There is no variation in technological
characteristics.
Test conducted per ASTM D3578, ASTM D512
indicates that the product meet the requirements.
Primary Skin irritation test ASTM F 719-81
Dermal sensitization Test ASTM F 720-81 (86) indicates no sensitization or irritation.
Not carried out
Non-clinical tests and biocompatibility tests
indicate device meet all performance and
biocompatibility requirements.
None



DEC 21 2000

Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

Mr. Chin-Guan Low Managing Director Terang Nusa SDN BHD 1 Jalan 8, Pengkalan Chepa 2 Industrial Zone 16100 Kota Bharu, Kelantan, MALAYSIA

Re: K003547

Trade Name: NUGARD DG Powdered Green-Mint Latex Examination Gloves With Protein Content Labeling

Claim (100 micrograms or less)

Regulatory Class: I Product Code: LYY

Dated: November 8, 2000 Received: November 17, 2000

Dear Mr. Low:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note:

this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4692. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "http://www.fda.gov/cdrh/dsma/dsmamain.html".

Ulatowski Timothy A.

Director

Division of Dental, Infection Control and General Hospital Devices Office of Device Evaluation Center for Devices and

Radiological Health

Enclosure



TERANG NUSA Sdn Bhd

510(k) Submission NUGARD DG

Indication for use Statement 3.

Submitter

Terang Nusa Sdn Bhd

510(k) Number

K003547

Device Name

Latex Examination Glove, Green-Mint, Towdered,

Contains 100 microgram or less of Water

Extractable Protein

Trade Name

NUGARD DG

Indication for use

These examination gloves are for use by healthcare workers or similar personnel during work to prevent cross contamination or cross infection between the user and the patient.

Concurrence of CDHR Office of Device Evaluation (ODE)

wision of Dental, Infection Control, eneral Hospital Devices 11 (19) Number <u>K 00 354</u>5

Prescription Use

OR

Over the counter